

International Independent Inspection & Testing Company

CERTIFICATE OF ANALYSIS

Client : Wathnan Industries Lab Report No. : GCD/LR/629/2019

Sample No. : GCD/ 629

Grade : Recycled Base Oil

Location : Qatar Seal No. : N/A

Date : 05/11/2019

| PARAMETER | METHOD | UNIT | RESULTS |
|-------------------------|-------------|--------|---------|
| Specific Gravity @ 60°F | ASTM D 1298 | - | 0.8591 |
| Flash Point | ASTM D 92 | °C | 160.0 |
| Pour Point | ASTM D 97 | °C | -9.0 |
| Viscosity @40 °C | ASTM D 445 | cST | 17.92 |
| Viscosity @100 °C | ASTM D 445 | cST | 3.801 |
| Viscosity Index | ASTM D 2270 | - | 101 |
| Acid No. | ASTM D 974 | mg/KOH | 0.07 |
| Color | ASTM D 1500 | - | <2.5 |

Test Conducted on 05/11/2019
Test Method Deviation None
Sampling Method N/A
Test Conducted by Emp# 696
Report Prepared by Emp# 696

The above test results are only applicable to the sample(s) referred above.



NISHAN SAGAR CHIEF CHEMIST

The Test Report shall not be reproduced (except in full) without the written approval of Geo –Chem Doha Samples retention period as per procedure QP/12, which shall be provided on request

QA: F/48; Revision: 01; Revision Date: 15/01/2019.
END OF REPORT